Figure 17


Correlation between Total PSA Assay (MST Pro Meter/MST Strip) and Total PSA Assay (Victor V/MST Strip)
y = 21.08x + 849.9
R² = 0.989
Fluorescence Intensity (Victor V)
25000
20000
15000
10000
5000
0
0
200
400
600
800
1000
ADC Counts (MST Pro Meter)