Figure 18


Correlation between Total PSA Assay (MST Pro Meter/MST strip) and Total PSA Assay (Victor V Microtiter Plate Assay)
y = 0.014x + 55.69
R² = 0.997
ADC Counts (MST Pro Meter/Strip Assay)
900
800
700
600
500
400
300
200
100
0
0
10000
20000
30000
40000
50000
60000
Fluorescence Intensity (Victor V - Microtiter Plate Assay)